

Certificado No.:

68422

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-04-17

ERM

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
10:15:00 AM	25.8	25.3	-0.50
10:16:00 AM	25.8	25.2	-0.60
10:17:00 AM	25.8	25.3	-0.50
10:18:00 AM	25.9	25.4	-0.50
10:19:00 AM	25.9	25.4	-0.50
10:20:00 AM	26.0	25.5	-0.50
10:21:00 AM	26.0	25.7	-0.30
10:22:00 AM	26.0	25.8	-0.20
10:23:00 AM	26.0	25.9	-0.10
10:24:00 AM	26.1	26.0	-0.10
10:25:00 AM	26.1	26.1	0.00
10:26:00 AM	26.2	26.4	0.20
10:27:00 AM	26.2	26.5	0.30
10:28:00 AM	26.2	26.7	0.50
10:29:00 AM	26.2	26.8	0.60
10:30:00 AM	26.3	27.0	0.70
10:31:00 AM	26.3	27.0	0.70
10:32:00 AM	26.3	27.1	0.80
10:33:00 AM	26.4	27.3	0.90
10:34:00 AM	26.4	27.4	1.00
10:35:00 AM	26.4	27.5	1.10
10:36:00 AM	26.5	27.6	1.10
10:37:00 AM	26.5	27.7	1.20
10:38:00 AM	26.6	27.7	1.10
10:39:00 AM	26.6	27.7	1.10
10:40:00 AM	26.6	27.6	1.00
10:41:00 AM	26.7	27.8	1.10
10:42:00 AM	26.7	27.7	1.00
10:43:00 AM	26.7	27.7	1.00
10:44:00 AM	26.7	27.8	1.10
Promedio	26.26	26.69	
Desviación	0.295	0.934	
Max	26.70	27.80	
Min	25.80	25.20	

Certificado No.:

68422

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-04-17

ERM

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
10:15:00 AM	39.5	39.0	-0.50
10:16:00 AM	39.5	39.0	-0.50
10:17:00 AM	39.1	39.0	-0.10
10:18:00 AM	38.7	39.0	0.30
10:19:00 AM	38.4	39.0	0.60
10:20:00 AM	38.2	37.6	-0.60
10:21:00 AM	38.0	37.6	-0.40
10:22:00 AM	37.7	37.4	-0.30
10:23:00 AM	37.2	36.8	-0.40
10:24:00 AM	37.1	36.5	-0.60
10:25:00 AM	37.0	36.8	-0.20
10:26:00 AM	36.8	37.0	0.20
10:27:00 AM	36.7	36.2	-0.50
10:28:00 AM	36.4	35.7	-0.70
10:29:00 AM	36.2	35.9	-0.30
10:30:00 AM	36.0	36.1	0.10
10:31:00 AM	35.7	35.2	-0.50
10:32:00 AM	35.5	35.2	-0.30
10:33:00 AM	35.1	35.4	0.30
10:34:00 AM	35.0	34.9	-0.10
10:35:00 AM	34.7	34.0	-0.70
10:36:00 AM	34.7	34.0	-0.70
10:37:00 AM	34.6	34.2	-0.40
10:38:00 AM	34.6	34.7	0.10
10:39:00 AM	34.5	34.0	-0.50
10:40:00 AM	34.5	34.4	-0.10
10:41:00 AM	34.4	34.0	-0.40
10:42:00 AM	34.4	34.1	-0.30
10:43:00 AM	34.4	34.1	-0.30
10:44:00 AM	34.3	34.1	-0.20
Promedio	36.30	36.03	
Desviación	1.720	1.780	
Max	39.50	39.00	
Min	34.30	34.00	

Certificado No.:

68422

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-04-17

ERM

