



Certificado No.:

56233

Termocupla

Fecha de Calibración:

2025-04-23

FULL GAUGE

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
10:30:00 AM	5.30	4.10	-1.20
10:30:30 AM	5.10	4.40	-0.70
10:31:00 AM	4.80	5.30	0.50
10:31:30 AM	4.60	5.30	0.70
10:32:00 AM	4.40	5.40	1.00
10:32:30 AM	4.50	5.40	0.90
10:33:00 AM	4.50	5.70	1.20
10:33:30 AM	4.60	5.70	1.10
10:34:00 AM	4.70	5.70	1.00
10:34:30 AM	4.80	5.80	1.00
10:35:00 AM	4.80	5.80	1.00
10:35:30 AM	4.90	5.80	0.90
10:36:00 AM	5.00	5.90	0.90
10:36:30 AM	5.10	5.90	0.80
10:37:00 AM	5.20	5.80	0.60
10:37:30 AM	5.10	6.10	1.00
10:38:00 AM	4.80	6.10	1.30
10:38:30 AM	4.50	6.10	1.60
10:39:00 AM	4.30	6.10	1.80
10:39:30 AM	4.30	6.10	1.80
10:40:00 AM	4.40	6.10	1.70
10:40:30 AM	4.50	6.10	1.60
10:41:00 AM	4.60	6.10	1.50
10:41:30 AM	4.60	6.20	1.60
10:42:00 AM	4.70	6.20	1.50
10:42:30 AM	4.90	3.00	-1.90
10:43:00 AM	6.10	3.10	-3.00
10:43:30 AM	7.00	3.10	-3.90
10:44:00 AM	6.90	3.10	-3.80
10:44:30 AM	6.50	3.10	-3.40
<b>Promedio</b>	<b>4.98</b>	<b>5.29</b>	
<b>Desviación</b>	<b>0.717</b>	<b>1.112</b>	
<b>Max</b>	<b>7.00</b>	<b>6.20</b>	
<b>Min</b>	<b>4.30</b>	<b>3.00</b>	

