

Certificado No.:

57349

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-05-28

MONNIT

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
11:30:00 AM	25.5	25.3	-0.19
11:31:00 AM	25.5	25.3	-0.19
11:32:00 AM	25.5	25.3	-0.19
11:33:00 AM	25.5	25.3	-0.19
11:34:00 AM	25.5	25.3	-0.19
11:35:00 AM	25.5	25.3	-0.19
11:36:00 AM	25.5	25.3	-0.19
11:37:00 AM	25.5	25.3	-0.19
11:38:00 AM	25.5	25.3	-0.19
11:39:00 AM	25.5	25.3	-0.19
11:40:00 AM	25.5	25.3	-0.19
11:41:00 AM	25.5	25.3	-0.19
11:42:00 AM	25.5	25.3	-0.19
11:43:00 AM	25.4	25.3	-0.09
11:44:00 AM	25.4	25.3	-0.09
11:45:00 AM	25.4	25.3	-0.09
11:46:00 AM	25.5	25.3	-0.19
11:47:00 AM	25.5	25.3	-0.19
11:48:00 AM	25.4	25.3	-0.09
11:49:00 AM	25.4	25.3	-0.09
11:50:00 AM	25.4	25.3	-0.09
11:51:00 AM	25.4	25.3	-0.09
11:52:00 AM	25.5	25.3	-0.19
11:53:00 AM	25.4	25.3	-0.09
11:54:00 AM	25.5	25.3	-0.19
11:55:00 AM	25.5	25.3	-0.19
11:56:00 AM	25.4	25.3	-0.09
11:57:00 AM	25.4	25.3	-0.09
11:58:00 AM	25.4	25.3	-0.09
11:59:00 AM	25.4	25.3	-0.09
Promedio	25.46	25.31	
Desviación	0.050	0.000	
Max	25.50	25.31	
Min	25.40	25.31	

Certificado No.:

57349

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-05-28

MONNIT

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
8:58:00 AM	49.2	49.46	0.26
8:59:00 AM	49.1	49.46	0.36
9:00:00 AM	49.1	49.46	0.36
9:01:00 AM	49.1	49.46	0.36
9:02:00 AM	49.1	49.46	0.36
9:03:00 AM	49.1	49.46	0.36
9:04:00 AM	49.1	49.46	0.36
9:05:00 AM	49.2	49.46	0.26
9:06:00 AM	49.2	49.46	0.26
9:07:00 AM	49.3	49.46	0.16
9:08:00 AM	49.3	49.46	0.16
9:09:00 AM	49.4	49.46	0.06
9:10:00 AM	49.4	49.46	0.06
9:11:00 AM	49.5	49.46	-0.04
9:12:00 AM	49.4	49.46	0.06
9:13:00 AM	49.6	49.46	-0.14
9:14:00 AM	49.5	49.46	-0.04
9:15:00 AM	49.5	49.46	-0.04
9:16:00 AM	49.5	49.46	-0.04
9:17:00 AM	49.6	49.46	-0.14
9:18:00 AM	49.7	49.46	-0.24
9:19:00 AM	49.7	49.46	-0.24
9:20:00 AM	49.7	49.46	-0.24
9:21:00 AM	49.8	49.46	-0.34
9:22:00 AM	49.8	49.46	-0.34
9:23:00 AM	49.8	49.46	-0.34
9:24:00 AM	49.8	49.46	-0.34
9:25:00 AM	49.9	49.46	-0.44
9:26:00 AM	49.9	49.46	-0.44
9:27:00 AM	50.0	54.87	4.87
Promedio	49.48	49.64	
Desviación	0.287	0.988	
Max	50.00	49.46	
Min	49.10	49.46	

