



Certificado No.:

66783

Equipo de Enfriamiento

Fecha de Calibración:

2026-03-03

THERMOKING

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
12:20:00 PM	2.7	3.1	0.40
12:21:00 PM	2.6	3.0	0.40
12:22:00 PM	2.5	2.9	0.40
12:23:00 PM	2.4	2.8	0.40
12:24:00 PM	2.3	2.6	0.30
12:25:00 PM	2.1	2.5	0.40
12:26:00 PM	2.1	2.4	0.30
12:27:00 PM	2.0	2.3	0.30
12:28:00 PM	1.8	2.2	0.40
12:29:00 PM	1.8	2.1	0.30
12:30:00 PM	1.7	2.0	0.30
12:31:00 PM	1.5	1.9	0.40
12:32:00 PM	1.4	1.8	0.40
12:33:00 PM	1.3	1.7	0.40
12:34:00 PM	1.2	1.5	0.30
12:35:00 PM	1.1	1.4	0.30
12:36:00 PM	1.0	1.3	0.30
12:37:00 PM	0.8	1.1	0.30
12:38:00 PM	0.8	1.0	0.20
12:39:00 PM	0.6	0.9	0.30
12:40:00 PM	0.6	0.8	0.20
12:41:00 PM	0.5	0.7	0.20
12:42:00 PM	0.4	0.6	0.20
12:43:00 PM	0.2	0.5	0.30
12:44:00 PM	0.2	0.4	0.20
12:45:00 PM	0.1	0.3	0.20
12:46:00 PM	-0.1	0.3	0.40
12:47:00 PM	-0.2	0.1	0.30
12:48:00 PM	-0.2	0.0	0.20
12:49:00 PM	-0.3	0.0	0.30
<b>Promedio</b>	<b>1.16</b>	<b>1.5</b>	
<b>Desviación</b>	<b>0.932</b>	<b>0.979</b>	
<b>Max</b>	<b>2.70</b>	<b>3.10</b>	
<b>Min</b>	<b>-0.30</b>	<b>0.00</b>	

Certificado No.:

66783

Equipo de Enfriamiento

Fecha de Calibración:

2026-03-03

THERMOKING

