

Certificado No.:

66784

Controlador de Temperatura

Fecha de Calibración:

2026-03-03

QUECLINCK

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
1:00:00 PM	-1.4	-0.9	0.50
1:01:00 PM	-1.6	-0.9	0.70
1:02:00 PM	-1.7	-0.9	0.80
1:03:00 PM	-1.9	-0.9	1.00
1:04:00 PM	-2.0	-0.9	1.10
1:05:00 PM	-2.1	-2.9	-0.80
1:06:00 PM	-2.2	-2.9	-0.70
1:07:00 PM	-2.4	-2.9	-0.50
1:08:00 PM	-2.7	-2.9	-0.20
1:09:00 PM	-2.8	-2.9	-0.10
1:10:00 PM	-3.0	-2.9	0.10
1:11:00 PM	-2.9	-2.9	0.00
1:12:00 PM	-3.2	-2.9	0.30
1:13:00 PM	-3.5	-2.9	0.60
1:14:00 PM	-3.6	-2.9	0.70
1:15:00 PM	-3.7	-4.8	-1.10
1:16:00 PM	-3.9	-4.8	-0.90
1:17:00 PM	-3.8	-4.8	-1.00
1:18:00 PM	-4.3	-4.8	-0.50
1:19:00 PM	-4.5	-4.8	-0.30
1:20:00 PM	-4.9	-4.8	0.10
1:21:00 PM	-4.9	-4.8	0.10
1:22:00 PM	-5.2	-4.8	0.40
1:23:00 PM	-5.4	-4.8	0.60
1:24:00 PM	-5.7	-4.8	0.90
1:25:00 PM	-5.6	-4.8	0.80
1:26:00 PM	-5.9	-6.4	-0.50
1:27:00 PM	-6.1	-6.4	-0.30
1:28:00 PM	-6.0	-6.4	-0.40
1:29:00 PM	-6.1	-6.4	-0.30
Promedio	-3.77	-3.7	
Desviación	1.529	1.742	
Max	-1.40	-0.90	
Min	-6.10	-6.40	

Certificado No.:

66784

Controlador de Temperatura

Fecha de Calibración:

2026-03-03

QUECLINCK

