

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.320 °C	-25.2 °C	0.120 °C	± 0.5 °C	± 4.0E-01 °C
3 °C	2.807 °C	3.0 °C	0.193 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	31.536 °C	31.6 °C	0.064 °C	± 0.5 °C	± 4.0E-01 °C