

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 29.942 °C                  | 29.8 °C            | -0.142 °C              | ± 1 °C                        | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.732 %HR                 | 31 %HR             | 1.268 %HR              | ± 5 %HR                       | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.642 %HR                 | 50 %HR             | -0.642 %HR             | ± 5 %HR                       | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.601 %HR                 | 73 %HR             | 2.399 %HR              | ± 5 %HR                       | ± 1.9E+00 %HR                        |