

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
30 °C	29.942 °C	29.8 °C	-0.142 °C	±	1 °C	± 4.0E-01 °C
30 %HR	29.732 %HR	31 %HR	1.268 %HR	±	5 %HR	± 1.9E+00 %HR
50 %HR	50.642 %HR	50 %HR	-0.642 %HR	±	5 %HR	± 1.9E+00 %HR
70 %HR	70.601 %HR	73 %HR	2.399 %HR	±	5 %HR	± 1.9E+00 %HR