

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.842 °C	29.8 °C	-0.042 °C	± 0.5 °C	± 1.6E-01 °C
36 °C	36.169 °C	36.1 °C	-0.069 °C	± 0.5 °C	± 1.6E-01 °C
37 °C	37.162 °C	37.1 °C	-0.062 °C	± 0.5 °C	± 1.6E-01 °C
40 °C	39.799 °C	39.9 °C	0.101 °C	± 0.5 °C	± 1.6E-01 °C