

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
23 °C	22.953 °C	23.3 °C	0.347 °C	± 0.8 °C	± 4.0E-01 °C
25 °C	25.018 °C	25.3 °C	0.282 °C	± 0.8 °C	± 4.0E-01 °C
30 °C	30.022 °C	29.7 °C	-0.322 °C	± 0.8 °C	± 4.0E-01 °C
36 °C	36.019 °C	36.4 °C	0.381 °C	± 0.8 °C	± 4.0E-01 °C
37 °C	37.272 °C	36.9 °C	-0.372 °C	± 0.8 °C	± 4.0E-01 °C
40 °C	39.889 °C	39.6 °C	-0.289 °C	± 0.8 °C	± 4.0E-01 °C