

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.872 °C	30.0 °C	0.128 °C	± 0.5 °C	± 1.60E-01 °C
24 %HR	22.779 %HR	25.2 %HR	2.421 %HR	± 5 %HR	± 1.40E+00 %HR
60 %HR	60.836 %HR	64.2 %HR	3.364 %HR	± 5.0 %HR	± 1.20E+00 %HR
90 %HR	90.202 %HR	93.8 %HR	3.598 %HR	± 5.0 %HR	± 1.10E+00 %HR

