

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
23 °C	22.953 °C	23.0 °C	0.047 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.018 °C	25.1 °C	0.082 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.022 °C	30.1 °C	0.078 °C	± 0.5 °C	± 4.0E-01 °C
36 °C	36.019 °C	36.2 °C	0.181 °C	± 0.5 °C	± 4.0E-01 °C
37 °C	37.272 °C	37.2 °C	-0.072 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.889 °C	40.1 °C	0.211 °C	± 0.5 °C	± 4.0E-01 °C