

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.842 °C	30.0 °C	0.158 °C	± 0.5 °C	± 4.0E-01 °C
36 °C	35.898 °C	36.0 °C	0.102 °C	± 0.5 °C	± 4.0E-01 °C
37 °C	36.921 °C	37.1 °C	0.179 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.919 °C	40.1 °C	0.181 °C	± 0.5 °C	± 4.0E-01 °C