

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
30 °C	29.842 °C	29.9 °C	0.058 °C	± 0.5 °C	± 4.0E-01 °C
36 °C	35.898 °C	36.0 °C	0.102 °C	± 0.5 °C	± 4.0E-01 °C
37 °C	36.921 °C	37.1 °C	0.179 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.919 °C	40.1 °C	0.181 °C	± 0.5 °C	± 4.0E-01 °C