

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.842 °C	29.9 °C	0.058 °C	± 0.5 °C	± 1.6E-01 °C
36 °C	35.898 °C	36.0 °C	0.102 °C	± 0.5 °C	± 1.6E-01 °C
37 °C	36.921 °C	37.1 °C	0.179 °C	± 0.5 °C	± 1.6E-01 °C
40 °C	39.919 °C	40.1 °C	0.181 °C	± 0.5 °C	± 1.6E-01 °C