

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.812 °C	29.8 °C	-0.012 °C	± 0.5 °C	± 1.60E-01 °C
36 °C	36.169 °C	35.2 °C	-0.969 °C	± 0.5 °C	± 1.40E+00 °C
37 °C	37.162 °C	36.4 °C	-0.762 °C	± 0.5 °C	± 1.20E+00 °C
40 °C	39.839 °C	38.9 °C	-0.939 °C	± 0.5 °C	± 1.10E+00 °C