

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.812 °C	29.8 °C	-0.012 °C	± 0.5 °C	± 1.60E-01 °C
36 °C	35.166 °C	35.2 °C	0.034 °C	± 0.5 °C	± 1.60E-01 °C
37 °C	37.162 °C	37.4 °C	0.238 °C	± 0.5 °C	± 1.60E-01 °C
40 °C	39.839 °C	39.8 °C	-0.039 °C	± 0.5 °C	± 1.60E-01 °C