

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
23.3 μm	23 μm	-0.3 μm	± 1.2 μm	±	5.8E-01 μm
51.6 μm	52 μm	0.4 μm	± 2.5 μm	±	5.8E-01 μm
124.6 μm	124 μm	-0.6 μm	± 3.2 μm	±	5.8E-01 μm
250.3 μm	254 μm	3.7 μm	± 4.5 μm	±	5.8E-01 μm
486.0 μm	494 μm	8.0 μm	± 6.9 μm	±	5.8E-01 μm
963.0 μm	964 μm	1.0 μm	± 12 μm	±	6.0E-01 μm