

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.931 °C	15.3 °C	0.369 °C	± 0.2 °C	± 4.0E-01 °C
25 °C	24.547 °C	24.7 °C	0.153 °C	± 0.2 °C	± 4.0E-01 °C
32 °C	32.108 °C	32.3 °C	0.192 °C	± 0.2 °C	± 4.0E-01 °C
70 %HR	68.210 %HR	70.0 %HR	1.790 %HR	± 5.0 %HR	± 1.9E+00 %HR