

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.084 °C	15.4 °C	0.316 °C	± 0.2 °C	± 4.0E-01 °C
25 °C	25.543 °C	25.7 °C	0.157 °C	± 0.2 °C	± 4.0E-01 °C
32 °C	32.149 °C	32.3 °C	0.151 °C	± 0.2 °C	± 4.0E-01 °C
70 %HR	69.860 %HR	71.0 %HR	1.140 %HR	± 5.0 %HR	± 1.9E+00 %HR