

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	14.740 °C	15.2 °C	0.460 °C	± 0.2 °C	± 4.0E-01 °C
25 °C	24.738 °C	24.5 °C	-0.238 °C	± 0.2 °C	± 4.0E-01 °C
32 °C	31.677 °C	31.4 °C	-0.277 °C	± 0.2 °C	± 4.0E-01 °C
70 %HR	69.425 %HR	68.0 %HR	-1.425 %HR	± 5.0 %HR	± 1.9E+00 %HR