

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.061 °C	14.7 °C	-0.361 °C	± 0.2 °C	± 4.0E-01 °C
25 °C	24.830 °C	25.0 °C	0.170 °C	± 0.2 °C	± 4.0E-01 °C
32 °C	31.919 °C	32.1 °C	0.181 °C	± 0.2 °C	± 4.0E-01 °C
70 %HR	69.200 %HR	68.0 %HR	-1.200 %HR	± 5.0 %HR	± 1.9E+00 %HR