

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.606 °C	15.0 °C	0.394 °C	± 0.2 °C	± 4.0E-01 °C
25 °C	24.758 °C	25.0 °C	0.242 °C	± 0.2 °C	± 4.0E-01 °C
32 °C	31.632 °C	31.4 °C	-0.232 °C	± 0.2 °C	± 4.0E-01 °C
70 %HR	68.587 %HR	70.5 %HR	1.913 %HR	± 5.0 %HR	± 1.9E+00 %HR