

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.290 °C	-23.9 °C	1.390 °C	± 0.8 °C	± 4.0E-01 °C
3 °C	2.988 °C	3.6 °C	0.612 °C	± 0.8 °C	± 4.0E-01 °C
32 °C	31.948 °C	31.8 °C	-0.148 °C	± 0.8 °C	± 4.0E-01 °C