

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.180 °C	0.5 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.466 °C	20.9 °C	0.434 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.297 °C	35.4 °C	0.103 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.672 %HR	34.6 %HR	3.928 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.315 %HR	54.6 %HR	-0.715 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.345 %HR	77.5 %HR	-2.845 %HR	± 3.0 %HR	± 1.9E+00 %HR