

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-18 °C	-18.271 °C	-17.9 °C	0.371 °C	± 1 °C	± 4.0E-01 °C
0 °C	-0.221 °C	-0.2 °C	0.021 °C	± 1 °C	± 4.0E-01 °C
4 °C	3.790 °C	3.8 °C	0.010 °C	± 1 °C	± 4.0E-01 °C
35 °C	35.136 °C	34.8 °C	-0.336 °C	± 1 °C	± 4.0E-01 °C
50 °C	49.897 °C	50.3 °C	0.403 °C	± 5 °C	± 4.0E-01 °C
30 %HR	29.271 %HR	30.0 %HR	0.729 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	71.041 %HR	75.0 %HR	3.959 %HR	± 5 %HR	± 1.9E+00 %HR