

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.310 °C	0.0 °C	-0.310 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.495 °C	5.0 °C	-0.495 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.0 °C	-0.138 °C	± 0.5 °C	± 4.0E-01 °C