

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.173 °C	30.4 °C	0.227 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	36.2 %HR	0.517 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.816 %HR	51.8 %HR	0.984 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.8 %HR	0.352 %HR	± 3.0 %HR	± 1.9E+00 %HR