

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.824 °C	31.1 °C	0.276 °C	± 1.0 °C	± 6.0E-01 °C
30 %HR	29.599 %HR	27 %HR	-2.599 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	50.396 %HR	50 %HR	-0.396 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	71.889 %HR	73 %HR	1.111 %HR	± 5 %HR	± 1.9E+00 %HR