

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.163 °C	30.2 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	32.0 %HR	-3.683 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.274 %HR	49.0 %HR	-1.274 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.754 %HR	66.0 %HR	-4.754 %HR	± 3.0 %HR	± 1.9E+00 %HR