

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.163 °C	29.9 °C	-0.263 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	26.0 %HR	-9.683 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.274 %HR	45.0 %HR	-5.274 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.754 %HR	65.0 %HR	-5.754 %HR	± 3.0 %HR	± 1.9E+00 %HR