

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
17 °C	17.327 °C	17.3 °C	-0.027 °C	± 0.5 °C	± 4.0E-01 °C
18 °C	18.450 °C	18.6 °C	0.150 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.193 °C	30.5 °C	0.307 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.196 °C	42.0 °C	-0.196 °C	± 0.5 °C	± 4.0E-01 °C
65 %HR	65.662 %HR	67.9 %HR	2.238 %HR	± 2.5 %HR	± 1.9E+00 %HR