

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.491 °C	-25.1 °C	0.391 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.000 °C	0.3 °C	0.300 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.4 °C	0.306 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.649 °C	10.9 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.968 °C	25.2 °C	0.232 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.461 °C	29.7 °C	0.239 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.110 °C	40.4 °C	0.290 °C	± 0.5 °C	± 4.0E-01 °C

