

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.491 °C	-24.9 °C	0.591 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.080 °C	0.4 °C	0.320 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.0 °C	-0.094 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.649 °C	10.7 °C	0.051 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	24.968 °C	25.2 °C	0.232 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	29.762 °C	30.1 °C	0.338 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.110 °C	40.2 °C	0.090 °C	± 0.45 °C	± 4.0E-01 °C

