

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.491 °C	-25.1 °C	0.391 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.080 °C	0.4 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.0 °C	-0.094 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.649 °C	10.7 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.968 °C	24.9 °C	-0.068 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.461 °C	29.6 °C	0.139 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.110 °C	40.2 °C	0.090 °C	± 0.5 °C	± 4.0E-01 °C

