

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.491 °C	-24.8 °C	0.691 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.000 °C	0.3 °C	0.300 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.2 °C	0.106 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.649 °C	10.4 °C	-0.249 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	24.968 °C	24.7 °C	-0.268 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	29.461 °C	29.1 °C	-0.361 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.391 °C	40.0 °C	-0.391 °C	± 0.45 °C	± 4.0E-01 °C

