

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.491 °C	-24.9 °C	0.591 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.000 °C	0.3 °C	0.300 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.294 °C	5.5 °C	0.206 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.649 °C	10.6 °C	-0.049 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	24.968 °C	25.2 °C	0.232 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	29.461 °C	29.5 °C	0.039 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.190 °C	39.9 °C	-0.290 °C	± 0.45 °C	± 4.0E-01 °C

