

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.491 °C	-25.0 °C	0.491 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.080 °C	0.4 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.3 °C	0.206 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.649 °C	11.0 °C	0.351 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.968 °C	25.1 °C	0.132 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.661 °C	30.0 °C	0.339 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.110 °C	40.1 °C	-0.010 °C	± 0.5 °C	± 4.0E-01 °C

