

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| -25 °C                     | -25.491 °C                 | -25.8 °C           | -0.309 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 0 °C                       | 0.000 °C                   | 0.3 °C             | 0.300 °C               | ± 0.45 °C                     | ± 4.0E-01 °C                         |
| 5 °C                       | 5.094 °C                   | 5.1 °C             | 0.006 °C               | ± 0.45 °C                     | ± 4.0E-01 °C                         |
| 10 °C                      | 10.649 °C                  | 10.4 °C            | -0.249 °C              | ± 0.45 °C                     | ± 4.0E-01 °C                         |
| 25 °C                      | 24.968 °C                  | 25.2 °C            | 0.232 °C               | ± 0.45 °C                     | ± 4.0E-01 °C                         |
| 30 °C                      | 29.461 °C                  | 29.8 °C            | 0.339 °C               | ± 0.45 °C                     | ± 4.0E-01 °C                         |
| 40 °C                      | 40.110 °C                  | 40.3 °C            | 0.190 °C               | ± 0.45 °C                     | ± 4.0E-01 °C                         |

