

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.491 °C	-25.8 °C	-0.309 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.000 °C	0.2 °C	0.200 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.094 °C	5.1 °C	0.006 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.649 °C	10.5 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.968 °C	25.2 °C	0.232 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.461 °C	29.7 °C	0.239 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.110 °C	40.3 °C	0.190 °C	± 0.5 °C	± 4.0E-01 °C

