

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.290 °C	-25.0 °C	0.290 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.210 °C	0.1 °C	-0.110 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.174 °C	5.4 °C	0.226 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.127 °C	10.1 °C	-0.027 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	19.754 °C	19.9 °C	0.146 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.042 °C	30.3 °C	0.258 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	39.699 °C	40.1 °C	0.401 °C	± 0.45 °C	± 4.0E-01 °C

