

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.290 °C	-25.6 °C	-0.310 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.210 °C	0.0 °C	-0.210 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.174 °C	5.1 °C	-0.074 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.127 °C	10.4 °C	0.273 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	19.754 °C	20.1 °C	0.346 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.042 °C	30.1 °C	0.058 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	39.699 °C	39.7 °C	0.001 °C	± 0.45 °C	± 4.0E-01 °C

