

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.7 °C	-0.410 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.210 °C	0.0 °C	-0.210 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.174 °C	5.1 °C	-0.074 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.127 °C	10.3 °C	0.173 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	19.754 °C	20.1 °C	0.346 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.042 °C	30.2 °C	0.158 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	39.699 °C	39.8 °C	0.101 °C	± 0.45 °C	± 4.0E-01 °C

