

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-20 °C	-20.076 °C	-19.9 °C	0.176 °C	± 0.2 °C	± 4.0E-01 °C
-10 °C	-10.379 °C	-10.2 °C	0.179 °C	± 0.2 °C	± 4.0E-01 °C
10 °C	9.967 °C	10.2 °C	0.193 °C	± 0.2 °C	± 4.0E-01 °C
20 °C	20.225 °C	20.4 °C	0.175 °C	± 0.2 °C	± 4.0E-01 °C
30 °C	30.263 °C	30.1 °C	-0.163 °C	± 0.2 °C	± 4.0E-01 °C
40 °C	39.990 °C	39.8 °C	-0.190 °C	± 0.2 °C	± 4.0E-01 °C
50 °C	49.987 °C	49.8 °C	-0.187 °C	± 0.2 °C	± 4.0E-01 °C
60 °C	59.985 °C	59.8 °C	-0.185 °C	± 0.2 °C	± 4.0E-01 °C
#10 %HR	10.273 %HR	12.2 %HR	1.927 %HR	± 2 %HR	± 1.9E+00 %HR
20 %HR	20.141 %HR	22.0 %HR	1.859 %HR	± 2 %HR	± 1.9E+00 %HR
30 %HR	30.621 %HR	32.6 %HR	1.979 %HR	± 2 %HR	± 1.9E+00 %HR
40 %HR	40.795 %HR	42.7 %HR	1.905 %HR	± 2 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	52.6 %HR	1.846 %HR	± 2 %HR	± 1.9E+00 %HR
60 %HR	60.368 %HR	58.9 %HR	1.468 %HR	± 2 %HR	± 1.9E+00 %HR
90 %HR	89.916 %HR	88.0 %HR	-1.916 %HR	± 2 %HR	± 1.9E+00 %HR