

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.2 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.5 %HR	1.930 %HR	± 2.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	52.6 %HR	1.917 %HR	± 2.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.7 %HR	1.252 %HR	± 2.0 %HR	± 1.9E+00 %HR