

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.280 °C	0.6 °C	0.320 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.975 °C	20.5 °C	0.525 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	29.972 °C	30.0 °C	0.028 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.098 %HR	31.5 %HR	2.402 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	52.175 %HR	53.4 %HR	1.225 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	77.155 %HR	79.5 %HR	2.345 %HR	± 3.0 %HR	± 1.9E+00 %HR