

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.280 °C	0.0 °C	-0.280 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.975 °C	19.6 °C	-0.375 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	29.972 °C	29.4 °C	-0.572 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.098 %HR	32.3 %HR	3.202 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	54.220 %HR	55.2 %HR	0.980 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	77.155 %HR	85.2 %HR	8.045 %HR	± 3.0 %HR	± 1.9E+00 %HR