

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.280 °C	0.1 °C	-0.180 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.975 °C	19.8 °C	-0.175 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	29.972 °C	29.5 °C	-0.472 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.098 %HR	32.5 %HR	3.402 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	54.220 %HR	53.8 %HR	-0.420 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	77.155 %HR	85.0 %HR	7.845 %HR	± 3.0 %HR	± 1.9E+00 %HR