

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.972 °C	30.1 °C	0.128 °C	± 0.5 °C	± 4.0E-01 °C
70 °C	70.072 °C	70.0 °C	-0.072 °C	± 0.5 °C	± 1.1E+00 °C
30 %HR	28.914 %HR	36.7 %HR	7.786 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.296 %HR	57.0 %HR	5.704 %HR	± 3.0 %HR	± 1.9E+00 %HR