

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	29.972 °C	30.1 °C	0.128 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.914 %HR	36.7 %HR	7.786 %HR	± 0.5 %HR	± 1.9E+00 %HR
50 %HR	51.296 %HR	57.0 %HR	5.704 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.527 %HR	70.0 %HR	0.473 %HR	± 3.0 %HR	± 1.9E+00 %HR